


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
Data File : VN053992.D
Acq On : 28 Feb 2019 9:03
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Title : SW846 8260
Last Update : Thu Feb 28 04:05:45 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	15769	PASS
75	95	30	60	47.9	39106	PASS
95	95	100	100	100.0	81578	PASS
96	95	5	9	6.4	5206	PASS
173	174	0.00	2	1.0	736	PASS
174	95	50	100	88.3	72045	PASS
175	174	5	9	6.9	4960	PASS
176	174	95	101	95.5	68816	PASS
177	176	5	9	6.7	4586	PASS