

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022820\
 Data File : VN060282.D
 Acq On : 28 Feb 2020 14:19
 Operator : JC/MD
 Sample : PB127123ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127123ZHE#13

Quant Time: Feb 29 01:20:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200636	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	326354	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308456	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127804	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
35) Dibromofluoromethane	7.57	113	89060	58.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.58%	
50) Toluene-d8	10.08	98	398624	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	12.40	95	152906	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
Target Compounds						
16) Acetone	3.79	43	9337	11.232	ug/l	91
20) Methylene Chloride	4.52	84	14804	6.485	ug/l	97
25) 2-Butanone	6.80	43	7717	5.747	ug/l	96
43) Isopropyl Acetate	8.14	43	54534	10.577	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022820\
Data File : VN060282.D
Acq On : 28 Feb 2020 14:19
Operator : JC/MD
Sample : PB127123ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127123ZHE#13

Quant Time: Feb 29 01:20:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

