

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022820\
 Data File : VN060283.D
 Acq On : 28 Feb 2020 14:41
 Operator : JC/MD
 Sample : PB127123ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127123ZHE#14

Quant Time: Feb 29 01:23:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	208508	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	340006	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	155111	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	134459	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
35) Dibromofluoromethane	7.57	113	90607	57.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.82%	
50) Toluene-d8	10.08	98	418216	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
62) 4-Bromofluorobenzene	12.40	95	160807	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
Target Compounds						
16) Acetone	3.79	43	8970	10.383	ug/l	100
20) Methylene Chloride	4.52	84	14268	6.014	ug/l	96
25) 2-Butanone	6.80	43	7672	5.498	ug/l	95
43) Isopropyl Acetate	8.14	43	53853	10.026	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022820\
Data File : VN060283.D
Acq On : 28 Feb 2020 14:41
Operator : JC/MD
Sample : PB127123ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127123ZHE#14

Quant Time: Feb 29 01:23:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

