

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022820\
 Data File : VN060295.D
 Acq On : 28 Feb 2020 19:13
 Operator : JC/MD
 Sample : L1717-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PLANK-MANHOLE

Quant Time: Feb 29 02:37:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	193157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	317207	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	297091	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144364	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	125161	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
35) Dibromofluoromethane	7.57	113	90033	61.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.30%	
50) Toluene-d8	10.08	98	387938	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
62) 4-Bromofluorobenzene	12.40	95	146602	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
Target Compounds						
16) Acetone	3.79	43	13972	17.459	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022820\
Data File : VN060295.D
Acq On : 28 Feb 2020 19:13
Operator : JC/MD
Sample : L1717-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PLANK-MANHOLE

Quant Time: Feb 29 02:37:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

