

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022822\
 Data File : VN071108.D
 Acq On : 28 Feb 2022 20:44
 Operator : JC\MD
 Sample : N1689-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-44

Quant Time: Mar 01 04:25:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	854136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1343852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1173004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	422570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	511670	52.442	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.880%	
35) Dibromofluoromethane	8.016	113	391636	48.996	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.000%	
50) Toluene-d8	10.439	98	1628711	50.738	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.480%	
62) 4-Bromofluorobenzene	12.727	95	558302	48.658	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.320%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	114728	64.521	ug/l #	90
16) Acetone	4.287	43	22038	4.101	ug/l	98
18) Methyl Acetate	4.863	43	6135	Below Cal		96
19) Methyl tert-butyl Ether	5.616	73	152203	5.189	ug/l	93
22) Diisopropyl ether	6.492	45	38930	1.273	ug/l #	77
25) 2-Butanone	7.345	43	4565	0.639	ug/l	99
95) Naphthalene	15.504	128	20246	5.954	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022822\
Data File : VN071108.D
Acq On : 28 Feb 2022 20:44
Operator : JC\MD
Sample : N1689-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-44

Quant Time: Mar 01 04:25:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

