

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076824.D
 Acq On : 28 Feb 2023 13:34
 Operator : JC\MD
 Sample : VN0228MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0228MBL01

Quant Time: Mar 01 01:25:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	279698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	497561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	439949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	168547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	199538	51.664	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.320%
35) Dibromofluoromethane	8.172	113	158503	48.644	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.280%
50) Toluene-d8	10.566	98	601442	49.425	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.854	95	179160	44.453	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076824.D
Acq On : 28 Feb 2023 13:34
Operator : JC\MD
Sample : VN0228MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0228MBL01

Quant Time: Mar 01 01:25:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

