

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076828.D
 Acq On : 28 Feb 2023 15:22
 Operator : JC\MD
 Sample : PB151081TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151081TB

Quant Time: Mar 01 01:28:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	286928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	506652	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	490403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	187317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	201180	50.777	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.560%
35) Dibromofluoromethane	8.172	113	155596	46.895	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	10.572	98	647209	52.232	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.460%
62) 4-Bromofluorobenzene	12.854	95	203452	49.574	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.140%
Target Compounds						
						Qvalue
16) Acetone	4.443	43	70791	39.999	ug/l	98
20) Methylene Chloride	5.290	84	20353	5.487	ug/l	93
37) Ethyl Acetate	7.566	43	62229	12.338	ug/l	99
43) Isopropyl Acetate	8.695	43	271503	33.080	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076828.D
Acq On : 28 Feb 2023 15:22
Operator : JC\MD
Sample : PB151081TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151081TB

Quant Time: Mar 01 01:28:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

