

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076830.D
 Acq On : 28 Feb 2023 16:10
 Operator : JC\MD
 Sample : PB151081ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151081ZHE#02

Quant Time: Mar 01 01:29:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	298840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	542896	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	527654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	192467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	214885	52.074	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.140%	
35) Dibromofluoromethane	8.172	113	164356	46.228	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	92.460%	
50) Toluene-d8	10.572	98	688816	51.878	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.760%	
62) 4-Bromofluorobenzene	12.854	95	214837	48.854	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	97.700%	
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	89695	48.661	ug/l	97
20) Methylene Chloride	5.290	84	30729	7.954	ug/l #	88
37) Ethyl Acetate	7.572	43	59494	11.008	ug/l	98
43) Isopropyl Acetate	8.695	43	337653	38.393	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076830.D
Acq On : 28 Feb 2023 16:10
Operator : JC\MD
Sample : PB151081ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151081ZHE#02

Quant Time: Mar 01 01:29:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

