

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076831.D
 Acq On : 28 Feb 2023 16:33
 Operator : JC\MD
 Sample : PB151081ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151081ZHE#03

Quant Time: Mar 01 01:29:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	286510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	524702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	507465	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	190169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	205596	51.967	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.940%
35) Dibromofluoromethane	8.172	113	157359	45.795	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.580%
50) Toluene-d8	10.572	98	655378	51.071	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.140%
62) 4-Bromofluorobenzene	12.854	95	201929	47.511	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.020%
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	88593	50.131	ug/l #	87
20) Methylene Chloride	5.290	84	27804	7.507	ug/l	89
37) Ethyl Acetate	7.566	43	56190	10.757	ug/l	96
43) Isopropyl Acetate	8.695	43	416319	48.979	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076831.D
Acq On : 28 Feb 2023 16:33
Operator : JC\MD
Sample : PB151081ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151081ZHE#03

Quant Time: Mar 01 01:29:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

