

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076832.D
 Acq On : 28 Feb 2023 16:57
 Operator : JC\MD
 Sample : PB151081ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB151081ZHE#04

Quant Time: Mar 01 01:30:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	285172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	517028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	507548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	191108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	198515	50.413	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.820%		
35) Dibromofluoromethane	8.172	113	154071	45.503	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 91.000%		
50) Toluene-d8	10.572	98	655988	51.878	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.760%		
62) 4-Bromofluorobenzene	12.854	95	202117	48.261	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.520%		
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	83784	47.632	ug/l #	87
20) Methylene Chloride	5.290	84	28036	7.605	ug/l	95
37) Ethyl Acetate	7.572	43	61363	11.922	ug/l #	95
43) Isopropyl Acetate	8.695	43	482663	57.628	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076832.D
Acq On : 28 Feb 2023 16:57
Operator : JC\MD
Sample : PB151081ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB151081ZHE#04

Quant Time: Mar 01 01:30:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

