

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076834.D
 Acq On : 28 Feb 2023 17:45
 Operator : JC\MD
 Sample : 01700-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1R

Quant Time: Mar 01 01:31:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	283704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	526293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	488104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	181091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	212892	54.343	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.680%	
35) Dibromofluoromethane	8.172	113	156192	45.318	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.640%	
50) Toluene-d8	10.572	98	669028	51.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.960%	
62) 4-Bromofluorobenzene	12.854	95	194043	45.517	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	76016	43.440	ug/l	95
20) Methylene Chloride	5.296	84	18736	5.109	ug/l	97
37) Ethyl Acetate	7.578	43	43958	8.390	ug/l	96
43) Isopropyl Acetate	8.695	43	396236	46.476	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076834.D
Acq On : 28 Feb 2023 17:45
Operator : JC\MD
Sample : 01700-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1R

Quant Time: Mar 01 01:31:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

