

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076837.D
 Acq On : 28 Feb 2023 18:57
 Operator : JC\MD
 Sample : 01724-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR53-VNJ201

Quant Time: Mar 01 01:32:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	307138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	577768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	531369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	196417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	230169	54.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.540%	
35) Dibromofluoromethane	8.172	113	172959	45.712	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 91.420%	
50) Toluene-d8	10.572	98	711363	50.343	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	12.854	95	214807	45.899	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.800%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	32197	16.995	ug/l	92
37) Ethyl Acetate	7.566	43	30982	5.386	ug/l #	85
43) Isopropyl Acetate	8.695	43	202137	21.597	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076837.D
Acq On : 28 Feb 2023 18:57
Operator : JC\MD
Sample : 01724-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR53-VNJ201

Quant Time: Mar 01 01:32:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

