

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
 Data File : VN076838.D
 Acq On : 28 Feb 2023 19:21
 Operator : JC\MD
 Sample : 01727-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JB-1S

Quant Time: Mar 01 01:33:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	262395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	491651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	479668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	176797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	190244	52.506	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.020%
35) Dibromofluoromethane	8.178	113	146648	45.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.100%
50) Toluene-d8	10.571	98	618718	51.456	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.920%
62) 4-Bromofluorobenzene	12.854	95	189515	47.587	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.180%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	88370	54.601	ug/l	98
20) Methylene Chloride	5.295	84	29847	8.799	ug/l #	91
37) Ethyl Acetate	7.572	43	56966	11.639	ug/l	97
43) Isopropyl Acetate	8.695	43	335974	42.184	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022823\
Data File : VN076838.D
Acq On : 28 Feb 2023 19:21
Operator : JC\MD
Sample : 01727-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JB-1S

Quant Time: Mar 01 01:33:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

