

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022824\
 Data File : VN081241.D
 Acq On : 28 Feb 2024 14:29
 Operator : JC\MD
 Sample : P1590-09DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D1-20240222DL

Quant Time: Feb 28 23:25:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	314491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	577361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	511385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	218070	56.571	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.140%			
35) Dibromofluoromethane	8.171	113	170138	52.113	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.220%			
50) Toluene-d8	10.571	98	647344	55.826	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.660%			
62) 4-Bromofluorobenzene	12.853	95	205837	47.242	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.480%			
Target Compounds						Qvalue
44) Trichloroethene	9.353	130	104948	24.338	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022824\
Data File : VN081241.D
Acq On : 28 Feb 2024 14:29
Operator : JC\MD
Sample : P1590-09DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D1-20240222DL

Quant Time: Feb 28 23:25:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

