

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022824\
 Data File : VN081243.D
 Acq On : 28 Feb 2024 15:17
 Operator : JC\MD
 Sample : P1590-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20240223

Quant Time: Feb 28 23:26:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	315688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	592187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	520769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	220379	56.954	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.900%
35) Dibromofluoromethane	8.171	113	176439	52.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.380%
50) Toluene-d8	10.571	98	664699	55.888	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.780%
62) 4-Bromofluorobenzene	12.853	95	211566	47.341	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.680%
Target Compounds						
52) Toluene	10.629	92	15447	1.469	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022824\
Data File : VN081243.D
Acq On : 28 Feb 2024 15:17
Operator : JC\MD
Sample : P1590-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20240223

Quant Time: Feb 28 23:26:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

