

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022824\
 Data File : VN081250.D
 Acq On : 28 Feb 2024 18:05
 Operator : JC\MD
 Sample : P1590-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB02-20240223

Quant Time: Feb 28 23:30:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	307147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	582461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	494306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	190088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	216085	57.397	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.800%
35) Dibromofluoromethane	8.171	113	170231	51.685	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.360%
50) Toluene-d8	10.571	98	641059	54.800	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.600%
62) 4-Bromofluorobenzene	12.853	95	202164	45.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.980%
Target Compounds						
16) Acetone	4.430	43	6835	4.614	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022824\
Data File : VN081250.D
Acq On : 28 Feb 2024 18:05
Operator : JC\MD
Sample : P1590-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB02-20240223

Quant Time: Feb 28 23:30:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

