

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
 Data File : VN081267.D
 Acq On : 29 Feb 2024 14:26
 Operator : JC\MD
 Sample : P1590-18DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D2-20240222DL

Quant Time: Feb 29 23:51:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	315157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	587717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	504767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	189052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	214838	55.615	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.240%			
35) Dibromofluoromethane	8.165	113	170445	51.287	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.580%			
50) Toluene-d8	10.571	98	653684	55.380	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.760%			
62) 4-Bromofluorobenzene	12.847	95	200640	45.237	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.480%			
Target Compounds						Qvalue
44) Trichloroethene	9.353	130	118349	26.962	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
Data File : VN081267.D
Acq On : 29 Feb 2024 14:26
Operator : JC\MD
Sample : P1590-18DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D2-20240222DL

Quant Time: Feb 29 23:51:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

