

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
 Data File : VN081268.D
 Acq On : 29 Feb 2024 14:50
 Operator : JC\MD
 Sample : P1590-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB02-20240223

Quant Time: Feb 29 23:52:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	334186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	617699	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	534881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	206466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	225783	55.120	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.240%
35) Dibromofluoromethane	8.165	113	175547	50.258	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.520%
50) Toluene-d8	10.565	98	679742	54.792	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.580%
62) 4-Bromofluorobenzene	12.853	95	216624	46.471	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%
Target Compounds						
16) Acetone	4.430	43	7012	4.351	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
Data File : VN081268.D
Acq On : 29 Feb 2024 14:50
Operator : JC\MD
Sample : P1590-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB02-20240223

Quant Time: Feb 29 23:52:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

