

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
 Data File : VN081271.D
 Acq On : 29 Feb 2024 17:27
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 29 23:53:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	64	0.00
2 T	Dichlorodifluoromethane	0.647	0.654	-1.1	61	0.00
3 P	Chloromethane	0.833	0.605	27.4#	54	0.00
4 C	Vinyl Chloride	0.718	0.675	6.0#	61	0.00
5 T	Bromomethane	0.510	0.432	15.3	57	0.00
6 T	Chloroethane	0.556	0.459	17.4	63	0.00
7 T	Trichlorofluoromethane	1.049	1.083	-3.2	67	0.00
8 T	Diethyl Ether	0.384	0.396	-3.1	66	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.585	0.628	-7.4	70	0.00
10 T	Methyl Iodide	0.653	0.603	7.7	57	0.00
11 T	Tert butyl alcohol	0.113	0.110	2.7	69	0.00
12 CM	1,1-Dichloroethene	0.565	0.565	0.0	65	0.00
13 T	Acrolein	0.060	0.039	35.0#	45#	0.00
14 T	Allyl chloride	0.839	0.821	2.1	64	0.00
15 T	Acrylonitrile	0.294	0.323	-9.9	73	0.00
16 T	Acetone	0.241	0.238	1.2	68	0.00
17 T	Carbon Disulfide	1.577	1.279	18.9	54	0.00
18 T	Methyl Acetate	0.588	0.757	-28.7#	82	0.00
19 T	Methyl tert-butyl Ether	1.937	2.072	-7.0	69	0.00
20 T	Methylene Chloride	0.693	0.677	2.3	70	0.00
21 T	trans-1,2-Dichloroethene	0.631	0.617	2.2	66	0.00
22 T	Diisopropyl ether	1.903	2.101	-10.4	71	0.00
23 T	Vinyl Acetate	1.544	1.631	-5.6	67	0.00
24 P	1,1-Dichloroethane	1.100	1.203	-9.4	71	0.00
25 T	2-Butanone	0.390	0.420	-7.7	73	0.00
26 T	2,2-Dichloropropane	1.031	1.100	-6.7	68	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.763	-4.1	68	0.00
28 T	Bromochloromethane	0.456	0.487	-6.8	70	0.00
29 T	Tetrahydrofuran	0.260	0.275	-5.8	72	0.00
30 C	Chloroform	1.170	1.309	-11.9#	73	0.00
31 T	Cyclohexane	1.058	0.979	7.5	65	0.00
32 T	1,1,1-Trichloroethane	1.065	1.133	-6.4	69	0.00
33 S	1,2-Dichloroethane-d4	0.613	0.658	-7.3	73	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	63	0.00
35 S	Dibromofluoromethane	0.283	0.329	-16.3	75	0.00
36 T	1,1-Dichloropropene	0.484	0.516	-6.6	69	0.00
37 T	Ethyl Acetate	0.437	0.467	-6.9	69	0.00
38 T	Carbon Tetrachloride	0.490	0.534	-9.0	67	0.00
39 T	Methylcyclohexane	0.565	0.584	-3.4	65	0.00
40 TM	Benzene	1.402	1.528	-9.0	69	0.00
41 T	Methacrylonitrile	0.249	0.274	-10.0	77	0.00
42 TM	1,2-Dichloroethane	0.489	0.545	-11.5	71	0.00
43 T	Isopropyl Acetate	0.765	0.844	-10.3	70	0.00
44 TM	Trichloroethene	0.373	0.388	-4.0	68	0.00
45 C	1,2-Dichloropropane	0.351	0.396	-12.8#	72	0.00
46 T	Dibromomethane	0.256	0.282	-10.2	72	0.00
47 T	Bromodichloromethane	0.484	0.565	-16.7	71	0.00
48 T	Methyl methacrylate	0.356	0.390	-9.6	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
 Data File : VN081271.D
 Acq On : 29 Feb 2024 17:27
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 29 23:53:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	70	0.00
50 S	Toluene-d8	1.004	1.041	-3.7	69	0.00
51 T	4-Methyl-2-Pentanone	0.440	0.519	-18.0	74	0.00
52 CM	Toluene	0.888	0.968	-9.0#	68	0.00
53 T	t-1,3-Dichloropropene	0.538	0.611	-13.6	67	0.00
54 T	cis-1,3-Dichloropropene	0.581	0.649	-11.7	69	0.00
55 T	1,1,2-Trichloroethane	0.341	0.393	-15.2	70	0.00
56 T	Ethyl methacrylate	0.522	0.591	-13.2	69	0.00
57 T	1,3-Dichloropropane	0.590	0.674	-14.2	72	0.00
58 T	2-Chloroethyl Vinyl ether	0.262	0.228	13.0	55	0.00
59 T	2-Hexanone	0.349	0.386	-10.6	66	0.00
60 T	Dibromochloromethane	0.363	0.416	-14.6	68	0.00
61 T	1,2-Dibromoethane	0.359	0.395	-10.0	69	0.00
62 S	4-Bromofluorobenzene	0.377	0.423	-12.2	74	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	64	0.00
64 T	Tetrachloroethene	0.341	0.386	-13.2	78	0.00
65 PM	Chlorobenzene	1.082	1.125	-4.0	68	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.420	-8.0	68	0.00
67 C	Ethyl Benzene	1.936	2.080	-7.4#	69	0.00
68 T	m/p-Xylenes	0.742	0.787	-6.1	67	0.00
69 T	o-Xylene	0.736	0.764	-3.8	66	0.00
70 T	Styrene	1.166	1.276	-9.4	68	0.00
71 P	Bromoform	0.247	0.289	-17.0	69	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
73 T	Isopropylbenzene	4.375	4.175	4.6	69	0.00
74 T	N-amyl acetate	1.849	1.717	7.1	67	0.00
75 P	1,1,2,2-Tetrachloroethane	1.230	1.255	-2.0	76	0.00
76 T	1,2,3-Trichloropropane	1.136	1.122	1.2	76	0.00
77 T	Bromobenzene	0.975	0.925	5.1	71	0.00
78 T	n-propylbenzene	5.036	5.099	-1.3	72	0.00
79 T	2-Chlorotoluene	3.089	2.952	4.4	72	0.00
80 T	1,3,5-Trimethylbenzene	3.547	3.556	-0.3	71	0.00
81 T	trans-1,4-Dichloro-2-butene	0.441	0.431	2.3	70	0.00
82 T	4-Chlorotoluene	3.020	2.967	1.8	72	0.00
83 T	tert-Butylbenzene	3.067	2.996	2.3	71	0.00
84 T	1,2,4-Trimethylbenzene	3.534	3.622	-2.5	73	0.00
85 T	sec-Butylbenzene	4.264	4.474	-4.9	74	0.00
86 T	p-Isopropyltoluene	3.542	3.701	-4.5	74	0.00
87 T	1,3-Dichlorobenzene	1.843	1.839	0.2	75	0.00
88 T	1,4-Dichlorobenzene	1.811	1.827	-0.9	75	0.00
89 T	n-Butylbenzene	3.085	3.409	-10.5	77	0.00
90 T	Hexachloroethane	0.585	0.638	-9.1	74	0.00
91 T	1,2-Dichlorobenzene	1.778	1.753	1.4	74	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.256	0.238	7.0	69	0.00
93 T	1,2,4-Trichlorobenzene	0.992	1.015	-2.3	73	0.00
94 T	Hexachlorobutadiene	0.370	0.392	-5.9	77	0.00
95 T	Naphthalene	3.634	3.560	2.0	71	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
Data File : VN081271.D
Acq On : 29 Feb 2024 17:27
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 29 23:53:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.984	1.018	-3.5	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5