



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022924\
Data File : VN081261.D
Acq On : 29 Feb 2024 11:30
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Title : SW846 8260
Last Update : Fri Feb 16 23:45:20 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.7	12223	PASS
75	95	30	60	54.0	33605	PASS
95	95	100	100	100.0	62181	PASS
96	95	5	9	5.9	3698	PASS
173	174	0.00	2	2.0	868	PASS
174	95	50	100	70.0	43539	PASS
175	174	5	9	7.9	3451	PASS
176	174	95	101	95.6	41640	PASS
177	176	5	9	6.6	2748	PASS

Average of 12.841 to 12.853 min.: VN081261.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.90	893	46.75	549	56.95	2504	69.00	762
37.00	3925	47.00	170	59.90	699	70.00	57
38.00	3080	47.30	270	60.95	3576	71.85	42
38.90	826	47.80	117	61.95	3471	72.95	315
39.10	574	48.00	376	62.95	3566	74.00	12044
41.10	57	49.00	2815	63.95	375	75.00	33605
43.85	776	50.00	12223	64.90	197	75.95	2766
44.05	58	51.00	4396	65.10	282	76.20	959
44.80	100	52.00	74	65.70	63	76.85	437
44.95	599	54.95	215	66.95	211	77.85	337
46.00	54	55.95	1423	67.95	7096	78.90	3349

Average of 12.841 to 12.853 min.: VN081261.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
79.85	822	92.95	3337	116.90	464	136.95	160
80.85	3276	93.95	7979	117.65	318	140.70	529
81.85	964	95.00	62181	118.60	99	141.05	614
83.00	56	96.00	3698	118.90	292	141.95	143
85.80	56	97.00	146	127.60	83	142.70	214
86.90	2551	103.60	81	127.80	105	142.95	1063
87.80	2851	104.00	308	128.85	115	144.50	55
90.80	335	105.70	59	129.60	176	144.95	242
91.00	265	105.90	347	129.90	155	145.95	202
91.70	313	112.90	50	131.00	81	171.90	580
91.95	1895	115.85	197	134.90	53	172.30	60

Average of 12.841 to 12.853 min.: VN081261.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
172.90	868						
173.95	43539						
174.95	3451						
175.90	41640						
176.95	2748						