

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030119\
 Data File : VN054022.D
 Acq On : 1 Mar 2019 11:57
 Operator : JC/SP
 Sample : VN0301WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0301WBL01

Quant Time: Mar 01 23:57:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N030119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 01 11:11:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	115910	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	624157	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	510463	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	242686	29.82	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.40%
60) 4-Bromofluorobenzene	12.40	95	191010	24.45	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	81.50%
63) Toluene-d8	10.09	98	743224	30.26	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.87%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030119\
Data File : VN054022.D
Acq On : 1 Mar 2019 11:57
Operator : JC/SP
Sample : VN0301WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0301WBL01

Quant Time: Mar 01 23:57:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N030119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 01 11:11:48 2019
Response via : Initial Calibration

