

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030119\
 Data File : VN054026.D
 Acq On : 1 Mar 2019 13:44
 Operator : JC/SP
 Sample : K1587-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-MH-C02

Quant Time: Mar 05 01:15:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N030119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 00:48:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	109525	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	570173	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	480098	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	237471	30.88	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.93%
60) 4-Bromofluorobenzene	12.40	95	175932	23.94	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	79.80%
63) Toluene-d8	10.09	98	692723	29.99	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.97%

Target Compounds

					Ovalue
15) Acetone	3.82	58	2466	3.498 ug/l	# 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030119\
Data File : VN054026.D
Acq On : 1 Mar 2019 13:44
Operator : JC/SP
Sample : K1587-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-MH-C02

Quant Time: Mar 05 01:15:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N030119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 00:48:50 2019
Response via : Initial Calibration

