

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
 Data File : VN066076.D
 Acq On : 1 Mar 2021 14:45
 Operator : JC/MD
 Sample : M1519-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW805-29-34-022621

Quant Time: Mar 01 21:42:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	119100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	190898	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	172926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	66301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	80267	48.19	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.38%
35) Dibromofluoromethane	7.553	113	58061	49.14	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.28%
50) Toluene-d8	10.058	98	232416	47.92	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.84%
62) 4-Bromofluorobenzene	12.373	95	82086	45.09	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.18%
Target Compounds						
					Qvalue	
16) Acetone	3.795	43	4716	7.46	ug/l	98
30) Chloroform	7.335	83	20433	7.84	ug/l	98
47) Bromodichloromethane	9.367	83	9553	5.21	ug/l	96
60) Dibromochloromethane	10.872	129	2556	1.88	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
Data File : VN066076.D
Acq On : 1 Mar 2021 14:45
Operator : JC/MD
Sample : M1519-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW805-29-34-022621

Quant Time: Mar 01 21:42:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

