

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
 Data File : VN066078.D
 Acq On : 1 Mar 2021 15:31
 Operator : JC/MD
 Sample : M1519-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW806-31-36-022621

Quant Time: Mar 01 21:43:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	113974	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	185602	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	169256	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	69473	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	76661	48.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.20%
35) Dibromofluoromethane	7.550	113	57235	49.83	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.66%
50) Toluene-d8	10.058	98	226650	48.07	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.14%
62) 4-Bromofluorobenzene	12.373	95	82990	46.88	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.76%

Target Compounds						Qvalue
4) Vinyl Chloride	2.184	62	716	0.39	ug/l #	89
12) 1,1-Dichloroethene	3.714	96	48290	37.73	ug/l	92
21) trans-1,2-Dichloroethene	5.004	96	6765	5.03	ug/l	95
24) 1,1-Dichloroethane	5.804	63	13765	5.45	ug/l	97
27) cis-1,2-Dichloroethene	6.788	96	79064	51.59	ug/l	94
40) Benzene	8.003	78	3849	0.69	ug/l	97
42) 1,2-Dichloroethane	8.090	62	10998	5.96	ug/l	92
44) Trichloroethene	8.801	130	19944	14.36	ug/l	97
49) 1,4-Dioxane	9.164	88	1324	58.23	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
Data File : VN066078.D
Acq On : 1 Mar 2021 15:31
Operator : JC/MD
Sample : M1519-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW806-31-36-022621

Quant Time: Mar 01 21:43:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

