

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
 Data File : VN066081.D
 Acq On : 1 Mar 2021 17:05
 Operator : JC/MD
 Sample : M1508-04DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1159-1160DL

Quant Time: Mar 01 22:26:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	114254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	184723	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	168910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	72512	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	76814	48.08	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.16%		
35) Dibromofluoromethane	7.553	113	56713	49.61	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.22%		
50) Toluene-d8	10.058	98	225428	48.04	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.08%		
62) 4-Bromofluorobenzene	12.373	95	83357	47.32	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.64%		
Target Compounds					Qvalue	
31) Cyclohexane	7.618	56	13851	5.81	ug/l	# 82
39) Methylcyclohexane	9.049	83	8997	4.40	ug/l	97
40) Benzene	8.004	78	227878	40.78	ug/l	100
52) Toluene	10.122	92	431589	125.75	ug/l	95
67) Ethyl Benzene	11.483	91	144940	22.67	ug/l	99
68) m/p-Xylenes	11.589	106	187179	78.46	ug/l	95
69) o-Xylene	11.917	106	86268	37.50	ug/l	96
73) Isopropylbenzene	12.222	105	10148	1.71	ug/l	99
78) n-propylbenzene	12.563	91	30803	4.58	ug/l	100
80) 1,3,5-Trimethylbenzene	12.704	105	34108	6.78	ug/l	97
84) 1,2,4-Trimethylbenzene	13.010	105	133772	26.51	ug/l	100
95) Naphthalene	15.100	128	24899	8.61	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
Data File : VN066081.D
Acq On : 1 Mar 2021 17:05
Operator : JC/MD
Sample : M1508-04DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1159-1160DL

Quant Time: Mar 01 22:26:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

