

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
 Data File : VN066085.D
 Acq On : 1 Mar 2021 18:40
 Operator : JC/MD
 Sample : M1488-06DL 4000X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41697DL

Quant Time: Mar 01 21:45:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	121606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	193130	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	176110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	73496	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	81543	47.95	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.90%		
35) Dibromofluoromethane	7.550	113	59127	49.47	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.94%		
50) Toluene-d8	10.058	98	233717	47.63	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.26%		
62) 4-Bromofluorobenzene	12.373	95	85806	46.59	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.18%		
Target Compounds						
						Qvalue
16) Acetone	3.788	43	345673	649.65	ug/l	94
25) 2-Butanone	6.795	43	3314	3.94	ug/l #	88
29) Tetrahydrofuran	7.174	42	8478	14.63	ug/l	95
52) Toluene	10.122	92	6783	1.89	ug/l	99
68) m/p-Xylenes	11.592	106	795	0.32	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
Data File : VN066085.D
Acq On : 1 Mar 2021 18:40
Operator : JC/MD
Sample : M1488-06DL 4000X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41697DL

Quant Time: Mar 01 21:45:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

