

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
 Data File : VN066088.D
 Acq On : 1 Mar 2021 19:51
 Operator : JC/MD
 Sample : M1488-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41700

Quant Time: Mar 01 21:46:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	112824	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	185309	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	169203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	73064	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	78567	49.80	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.60%
35) Dibromofluoromethane	7.553	113	56659	49.40	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.80%
50) Toluene-d8	10.058	98	225260	47.85	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.70%
62) 4-Bromofluorobenzene	12.373	95	83521	47.26	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.52%
Target Compounds						
					Qvalue	
16) Acetone	3.795	43	20812	40.65	ug/l	91
20) Methylene Chloride	4.521	84	8433	5.76	ug/l	98
25) 2-Butanone	6.798	43	4136	5.30	ug/l	92
43) Isopropyl Acetate	8.126	43	8619	3.28	ug/l #	90
52) Toluene	10.122	92	5775	1.68	ug/l	99
95) Naphthalene	15.100	128	6301	4.45	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030121\
Data File : VN066088.D
Acq On : 1 Mar 2021 19:51
Operator : JC/MD
Sample : M1488-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41700

Quant Time: Mar 01 21:46:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

