

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030121\  
 Data File : VN066089.D  
 Acq On : 1 Mar 2021 20:15  
 Operator : JC/MD  
 Sample : M1488-03  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 41701

Quant Time: Mar 01 21:47:24 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Feb 17 15:48:13 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 7.627  | 168   | 118634   | 50.00    | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.550  | 114   | 192590   | 50.00    | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.376 | 117   | 175651   | 50.00    | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.315 | 152   | 78543    | 50.00    | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 7.987  | 65    | 81277    | 48.99    | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =      | 97.98%   |
| 35) Dibromofluoromethane    | 7.553  | 113   | 59384    | 49.82    | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =      | 99.64%   |
| 50) Toluene-d8              | 10.058 | 98    | 238370   | 48.72    | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =      | 97.44%   |
| 62) 4-Bromofluorobenzene    | 12.373 | 95    | 87471    | 47.62    | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =      | 95.24%   |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 3.791  | 43    | 14964    | 27.28    | ug/l   | 98       |
| 20) Methylene Chloride      | 4.525  | 84    | 10248    | 6.66     | ug/l   | 97       |
| 40) Benzene                 | 8.003  | 78    | 22770    | 3.91     | ug/l   | 95       |
| 43) Isopropyl Acetate       | 8.126  | 43    | 7375     | 2.70     | ug/l # | 91       |
| 52) Toluene                 | 10.122 | 92    | 10968    | 3.07     | ug/l   | 94       |
| 95) Naphthalene             | 15.100 | 128   | 7548     | 4.61     | ug/l   | 98       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030121\  
Data File : VN066089.D  
Acq On : 1 Mar 2021 20:15  
Operator : JC/MD  
Sample : M1488-03  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
41701

Quant Time: Mar 01 21:47:24 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021721W.M  
Quant Title : SW846 8260  
QLast Update : Wed Feb 17 15:48:13 2021  
Response via : Initial Calibration

