

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
 Data File : VN071119.D
 Acq On : 01 Mar 2022 15:24
 Operator : JC\MD
 Sample : N1689-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Mar 02 03:29:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	944665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1506785	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1336631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	461313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	570219	52.842	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.680%
35) Dibromofluoromethane	8.022	113	434454	48.475	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.960%
50) Toluene-d8	10.439	98	1822380	50.633	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.727	95	619558	48.158	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.320%
Target Compounds						Qvalue
16) Acetone	4.310	43	12179	2.049	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
Data File : VN071119.D
Acq On : 01 Mar 2022 15:24
Operator : JC\MD
Sample : N1689-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Mar 02 03:29:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

