

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
 Data File : VN071125.D
 Acq On : 01 Mar 2022 17:48
 Operator : JC\MD
 Sample : N1689-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-41

Quant Time: Mar 02 03:31:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	858758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1374189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1221010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	407801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	523335	53.348	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.700%	
35) Dibromofluoromethane	8.022	113	400406	48.987	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.980%	
50) Toluene-d8	10.439	98	1669087	50.848	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.700%	
62) 4-Bromofluorobenzene	12.727	95	552119	47.057	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.120%	
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	16354	3.027	ug/l	94
95) Naphthalene	15.498	128	9139	5.441	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
Data File : VN071125.D
Acq On : 01 Mar 2022 17:48
Operator : JC\MD
Sample : N1689-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-41

Quant Time: Mar 02 03:31:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

