

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
 Data File : VN071128.D
 Acq On : 01 Mar 2022 19:00
 Operator : JC\MD
 Sample : N1689-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-44

Quant Time: Mar 02 03:32:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	826749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1326919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1202625	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	410383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	513162	54.337	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.680%
35) Dibromofluoromethane	8.016	113	384522	48.720	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.440%
50) Toluene-d8	10.439	98	1632275	51.498	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.000%
62) 4-Bromofluorobenzene	12.727	95	563491	49.737	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.480%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	118711	68.973	ug/l	100
16) Acetone	4.299	43	23044	4.431	ug/l	90
19) Methyl tert-butyl Ether	5.610	73	147602	5.199	ug/l	95
22) Diisopropyl ether	6.498	45	39256	1.326	ug/l #	84
95) Naphthalene	15.504	128	23677	6.151	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
Data File : VN071128.D
Acq On : 01 Mar 2022 19:00
Operator : JC\MD
Sample : N1689-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-44

Quant Time: Mar 02 03:32:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

