

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
 Data File : VN071131.D
 Acq On : 01 Mar 2022 20:12
 Operator : JC\MD
 Sample : N1713-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-235

Quant Time: Mar 02 04:05:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	822282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1323326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1183313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	456501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	504207	53.679	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.360%	
35) Dibromofluoromethane	8.016	113	382983	48.657	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.320%	
50) Toluene-d8	10.439	98	1629929	51.564	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.120%	
62) 4-Bromofluorobenzene	12.727	95	592753	52.462	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.920%	
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	494896	95.670	ug/l	96
20) Methylene Chloride	5.092	84	17737	1.887	ug/l	92
43) Isopropyl Acetate	8.557	43	105106	4.822	ug/l #	82
69) o-Xylene	12.274	106	21086	1.252	ug/l	87
84) 1,2,4-Trimethylbenzene	13.362	105	123592	4.218	ug/l	92
86) p-Isopropyltoluene	13.610	119	39737	1.301	ug/l #	79
95) Naphthalene	15.498	128	22817	5.996	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030122\
Data File : VN071131.D
Acq On : 01 Mar 2022 20:12
Operator : JC\MD
Sample : N1713-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-235

Quant Time: Mar 02 04:05:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

