

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081274.D
 Acq On : 01 Mar 2024 10:25
 Operator : JC\MD
 Sample : VN0301MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0301MBL01

Quant Time: Mar 01 22:12:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	288998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	536728	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	454976	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	178516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	214327	60.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.000%	
35) Dibromofluoromethane	8.165	113	162737	53.620	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.240%	
50) Toluene-d8	10.565	98	637470	59.137	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 118.280%#	
62) 4-Bromofluorobenzene	12.847	95	189915	46.887	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081274.D
Acq On : 01 Mar 2024 10:25
Operator : JC\MD
Sample : VN0301MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0301MBL01

Quant Time: Mar 01 22:12:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

