

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081280.D
 Acq On : 01 Mar 2024 13:08
 Operator : JC\MD
 Sample : PB153949ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153949ZHE#02

Quant Time: Mar 01 22:16:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	311213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	589258	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	519007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	230114	60.325	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.640%	
35) Dibromofluoromethane	8.165	113	185026	55.529	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.060%	
50) Toluene-d8	10.571	98	713022	60.249	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 120.500%#	
62) 4-Bromofluorobenzene	12.847	95	216241	48.627	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.260%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	70883	47.225	ug/l	92
18) Methyl Acetate	5.018	43	4171	1.139	ug/l	91
20) Methylene Chloride	5.271	84	31154	7.228	ug/l	93
37) Ethyl Acetate	7.565	43	5939	1.153	ug/l	97
43) Isopropyl Acetate	8.688	43	434103	48.177	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081280.D
Acq On : 01 Mar 2024 13:08
Operator : JC\MD
Sample : PB153949ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153949ZHE#02

Quant Time: Mar 01 22:16:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

