

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081281.D
 Acq On : 01 Mar 2024 13:32
 Operator : JC\MD
 Sample : PB153949ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153949ZHE#03

Quant Time: Mar 01 22:16:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	312217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	588434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	529122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	205451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	226076	59.075	ug/l	0.00
Spiked Amount 50.000	Range	74 - 125	Recovery	=	118.160%	
35) Dibromofluoromethane	8.165	113	184363	55.407	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	110.820%	
50) Toluene-d8	10.565	98	718087	60.762	ug/l	0.00
Spiked Amount 50.000	Range	86 - 113	Recovery	=	121.520%#	
62) 4-Bromofluorobenzene	12.847	95	226843	51.083	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	102.160%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	72907	48.417	ug/l	94
18) Methyl Acetate	5.030	43	4247	1.156	ug/l #	76
20) Methylene Chloride	5.277	84	30569	7.049	ug/l	88
37) Ethyl Acetate	7.553	43	6029	1.172	ug/l #	94
43) Isopropyl Acetate	8.688	43	435948	48.450	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081281.D
Acq On : 01 Mar 2024 13:32
Operator : JC\MD
Sample : PB153949ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153949ZHE#03

Quant Time: Mar 01 22:16:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

