

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081282.D
 Acq On : 01 Mar 2024 13:56
 Operator : JC\MD
 Sample : PB153949ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153949ZHE#04

Quant Time: Mar 01 22:17:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	310766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	585644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	527408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	200043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	230578	60.533	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 121.060%		
35) Dibromofluoromethane	8.165	113	177052	53.464	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.920%		
50) Toluene-d8	10.565	98	710448	60.402	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 120.800%#		
62) 4-Bromofluorobenzene	12.853	95	224404	50.774	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.540%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	70204	46.840	ug/l	96
20) Methylene Chloride	5.283	84	30287	7.012	ug/l	87
37) Ethyl Acetate	7.559	43	7009	1.369	ug/l	99
43) Isopropyl Acetate	8.688	43	432016	48.241	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081282.D
Acq On : 01 Mar 2024 13:56
Operator : JC\MD
Sample : PB153949ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153949ZHE#04

Quant Time: Mar 01 22:17:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

