

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081286.D
 Acq On : 01 Mar 2024 15:31
 Operator : JC\MD
 Sample : P1602-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Mar 01 22:19:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	328735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	627394	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	564119	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	218192	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	229134	56.866	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.740%	
35) Dibromofluoromethane	8.171	113	178293	50.256	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.520%	
50) Toluene-d8	10.565	98	711987	56.504	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.000%	
62) 4-Bromofluorobenzene	12.847	95	217912	46.024	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	75462	47.596	ug/l	100
18) Methyl Acetate	5.024	43	4859	1.256	ug/l #	89
20) Methylene Chloride	5.277	84	27827	5.966	ug/l #	81
37) Ethyl Acetate	7.571	43	6013	1.096	ug/l #	92
43) Isopropyl Acetate	8.688	43	445463	46.433	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081286.D
Acq On : 01 Mar 2024 15:31
Operator : JC\MD
Sample : P1602-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Mar 01 22:19:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

