

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081287.D
 Acq On : 01 Mar 2024 15:55
 Operator : JC\MD
 Sample : P1602-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-3

Quant Time: Mar 01 22:19:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	330062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	637401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	570301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	230389	56.948	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.900%	
35) Dibromofluoromethane	8.165	113	180237	50.006	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.020%	
50) Toluene-d8	10.570	98	710566	55.506	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	111.020%	
62) 4-Bromofluorobenzene	12.853	95	223810	46.528	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	93.060%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	140223	88.087	ug/l	97
18) Methyl Acetate	5.030	43	5017	1.292	ug/l #	83
20) Methylene Chloride	5.277	84	24823	5.195	ug/l #	79
43) Isopropyl Acetate	8.688	43	287945	29.543	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081287.D
Acq On : 01 Mar 2024 15:55
Operator : JC\MD
Sample : P1602-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-3

Quant Time: Mar 01 22:19:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

