

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
 Data File : VN081288.D
 Acq On : 01 Mar 2024 16:19
 Operator : JC\MD
 Sample : P1602-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Mar 01 22:20:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	318856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	612694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	537563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201158	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223443	57.172	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.340%
35) Dibromofluoromethane	8.165	113	171822	49.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.180%
50) Toluene-d8	10.571	98	678929	55.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.340%
62) 4-Bromofluorobenzene	12.853	95	208467	45.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.180%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	103299	67.172	ug/l	99
20) Methylene Chloride	5.283	84	23071	4.962	ug/l #	81
43) Isopropyl Acetate	8.688	43	222190	23.716	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030124\
Data File : VN081288.D
Acq On : 01 Mar 2024 16:19
Operator : JC\MD
Sample : P1602-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Mar 01 22:20:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

