

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030219\
 Data File : VN054043.D
 Acq On : 1 Mar 2019 21:29
 Operator : JC/SP
 Sample : K1675-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z-3-11-A

Manual Integrations
 APPROVED

MMDadoda
 3/4/2019 1:53:53 PM

Quant Time: Mar 02 05:44:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	623216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1142341	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1113418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	409505	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	404782	55.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.82%	
35) Dibromofluoromethane	7.59	113	360269	48.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.54%	
50) Toluene-d8	10.09	98	1449060	52.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.40	95	489757	53.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.92%	
Target Compounds						
20) Methylene Chloride	4.55	84	39600	5.477	ug/l	90
43) Isopropyl Acetate	8.19	43	44711m	3.491	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030219\
Data File : VN054043.D
Acq On : 1 Mar 2019 21:29
Operator : JC/SP
Sample : K1675-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z-3-11-A

Manual Integrations
APPROVED

MMDadoda
3/4/2019 1:53:53 PM

Quant Time: Mar 02 05:44:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

