

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030220\
 Data File : VN060307.D
 Acq On : 2 Mar 2020 13:34
 Operator : JC/MD
 Sample : PB127166ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127166ZHE#02

Quant Time: Mar 03 06:15:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	212873	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	345915	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	319977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	151588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	134495	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
35) Dibromofluoromethane	7.57	113	93774	58.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.80%	
50) Toluene-d8	10.08	98	427846	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
62) 4-Bromofluorobenzene	12.40	95	158611	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
16) Acetone	3.79	43	7299	8.276	ug/l	98
20) Methylene Chloride	4.52	84	2947	1.217	ug/l #	86
43) Isopropyl Acetate	8.14	43	68635	12.559	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030220\
Data File : VN060307.D
Acq On : 2 Mar 2020 13:34
Operator : JC/MD
Sample : PB127166ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127166ZHE#02

Quant Time: Mar 03 06:15:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

