

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030220\
 Data File : VN060308.D
 Acq On : 2 Mar 2020 13:56
 Operator : JC/MD
 Sample : PB127166ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127166ZHE#03

Quant Time: Mar 03 06:18:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	208880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	338454	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	317267	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	151894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	132375	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
35) Dibromofluoromethane	7.56	113	91115	58.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.00%	
50) Toluene-d8	10.08	98	413468	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	12.40	95	158183	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
Target Compounds						
16) Acetone	3.79	43	7006	8.095	ug/l	95
20) Methylene Chloride	4.53	84	2968	1.249	ug/l	91
43) Isopropyl Acetate	8.14	43	68105	12.737	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030220\
Data File : VN060308.D
Acq On : 2 Mar 2020 13:56
Operator : JC/MD
Sample : PB127166ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127166ZHE#03

Quant Time: Mar 03 06:18:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

