

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030220\
 Data File : VN060309.D
 Acq On : 2 Mar 2020 14:19
 Operator : JC/MD
 Sample : PB127166ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127166ZHE#04

Manual Integrations
 APPROVED

MMDadoda
 3/3/2020 10:04:08 AM

Quant Time: Mar 03 06:20:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	330173	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	307989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130112	50.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.02%	
35) Dibromofluoromethane	7.57	113	89299	58.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.54%	
50) Toluene-d8	10.08	98	408466	52.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.32%	
62) 4-Bromofluorobenzene	12.40	95	151044	51.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.74%	
Target Compounds						
16) Acetone	3.79	43	7084	8.432	ug/l	94
20) Methylene Chloride	4.52	84	2645m	1.146	ug/l	
43) Isopropyl Acetate	8.14	43	66547	12.758	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030220\
Data File : VN060309.D
Acq On : 2 Mar 2020 14:19
Operator : JC/MD
Sample : PB127166ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127166ZHE#04

Manual Integrations
APPROVED

MMDadoda
3/3/2020 10:04:08 AM

Quant Time: Mar 03 06:20:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

