

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030220\
 Data File : VN060311.D
 Acq On : 2 Mar 2020 15:05
 Operator : JC/MD
 Sample : L1372-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-022820

Quant Time: Mar 03 06:30:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195328	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	313744	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	291439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124445	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	7.57	113	86831	59.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.24%	
50) Toluene-d8	10.08	98	383781	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	12.40	95	143035	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
Target Compounds						
16) Acetone	3.79	43	9411	11.629	ug/l	89
20) Methylene Chloride	4.52	84	3615	1.626	ug/l	91
43) Isopropyl Acetate	8.14	43	54957	11.088	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030220\
Data File : VN060311.D
Acq On : 2 Mar 2020 15:05
Operator : JC/MD
Sample : L1372-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-022820

Quant Time: Mar 03 06:30:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

