

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030221\
 Data File : VN066109.D
 Acq On : 2 Mar 2021 19:26
 Operator : JC/MD
 Sample : M1501-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R-LA-3-7-WC

Quant Time: Mar 03 03:55:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 02 14:05:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	76672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	141473	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	149799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	61332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	57934	49.65	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.30%
35) Dibromofluoromethane	7.547	113	44014	49.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.26%
50) Toluene-d8	10.055	98	181348	51.97	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.94%
62) 4-Bromofluorobenzene	12.370	95	76285	56.76	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.52%
Target Compounds						
16) Acetone	3.785	43	45446	130.31	ug/l	96
20) Methylene Chloride	4.515	84	8193	8.92	ug/l	95
25) 2-Butanone	6.795	43	4955	10.71	ug/l	96
43) Isopropyl Acetate	8.122	43	3522	1.82	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030221\
Data File : VN066109.D
Acq On : 2 Mar 2021 19:26
Operator : JC/MD
Sample : M1501-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R-LA-3-7-WC

Quant Time: Mar 03 03:55:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 02 14:05:08 2021
Response via : Initial Calibration

