

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030221\
 Data File : VN066112.D
 Acq On : 2 Mar 2021 20:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 03 03:56:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 02 14:05:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	50.000	49.962	0.1	96	0.00
3 P	Chloromethane	50.000	48.294	3.4	97	0.00
4 C	Vinyl Chloride	50.000	49.052	1.9#	97	0.00
5 T	Bromomethane	50.000	47.503	5.0	97	0.00
6 T	Chloroethane	50.000	52.452	-4.9	101	0.00
7 T	Trichlorofluoromethane	50.000	50.411	-0.8	99	0.00
8 T	Diethyl Ether	50.000	50.407	-0.8	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.256	-0.5	96	0.00
10 T	Methyl Iodide	50.000	51.955	-3.9	98	0.00
11 T	Tert butyl alcohol	250.000	204.142	18.3	80	0.00
12 CM	1,1-Dichloroethene	50.000	47.511	5.0#	94	0.00
13 T	Acrolein	250.000	218.226	12.7	89	0.00
14 T	Allyl chloride	50.000	52.457	-4.9	97	0.00
15 T	Acrylonitrile	250.000	241.212	3.5	91	0.00
16 T	Acetone	250.000	206.351	17.5	87	0.00
17 T	Carbon Disulfide	50.000	47.866	4.3	97	0.00
18 T	Methyl Acetate	50.000	47.776	4.4	91	0.00
19 T	Methyl tert-butyl Ether	50.000	46.978	6.0	94	0.00
20 T	Methylene Chloride	50.000	47.626	4.7	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.386	-2.8	99	0.00
22 T	Diisopropyl ether	50.000	50.003	-0.0	97	0.00
23 T	Vinyl Acetate	250.000	245.996	1.6	95	0.00
24 P	1,1-Dichloroethane	50.000	50.962	-1.9	98	0.00
25 T	2-Butanone	250.000	223.089	10.8	86	0.00
26 T	2,2-Dichloropropane	50.000	48.051	3.9	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.013	-0.0	98	0.00
28 T	Bromochloromethane	50.000	45.877	8.2	94	0.00
29 T	Tetrahydrofuran	250.000	227.767	8.9	88	0.00
30 C	Chloroform	50.000	49.729	0.5#	99	0.00
31 T	Cyclohexane	50.000	47.998	4.0	100	0.00
32 T	1,1,1-Trichloroethane	50.000	49.892	0.2	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.471	1.1	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	48.756	2.5	98	0.00
36 T	1,1-Dichloropropene	50.000	50.509	-1.0	97	0.00
37 T	Ethyl Acetate	50.000	48.142	3.7	89	0.00
38 T	Carbon Tetrachloride	50.000	50.087	-0.2	98	0.00
39 T	Methylcyclohexane	50.000	50.107	-0.2	97	0.00
40 TM	Benzene	50.000	49.537	0.9	97	0.00
41 T	Methacrylonitrile	50.000	43.710	12.6	84	0.00
42 TM	1,2-Dichloroethane	50.000	50.806	-1.6	99	0.00
43 T	Isopropyl Acetate	50.000	48.319	3.4	95	0.00
44 TM	Trichloroethene	50.000	48.902	2.2	97	0.00
45 C	1,2-Dichloropropane	50.000	48.248	3.5#	96	0.00
46 T	Dibromomethane	50.000	52.050	-4.1	97	0.00
47 T	Bromodichloromethane	50.000	50.373	-0.7	98	0.00
48 T	Methyl methacrylate	50.000	46.937	6.1	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030221\
 Data File : VN066112.D
 Acq On : 2 Mar 2021 20:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 03 03:56:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 02 14:05:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	780.071	22.0	79	0.00
50 S	Toluene-d8	50.000	48.259	3.5	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	232.846	6.9	91	0.00
52 CM	Toluene	50.000	50.042	-0.1#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	48.863	2.3	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.697	2.6	95	0.00
55 T	1,1,2-Trichloroethane	50.000	48.756	2.5	98	0.00
56 T	Ethyl methacrylate	50.000	49.830	0.3	95	0.00
57 T	1,3-Dichloropropane	50.000	49.176	1.6	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	252.709	-1.1	95	0.00
59 T	2-Hexanone	250.000	235.027	6.0	88	0.00
60 T	Dibromochloromethane	50.000	50.605	-1.2	98	0.00
61 T	1,2-Dibromoethane	50.000	48.990	2.0	93	0.00
62 S	4-Bromofluorobenzene	50.000	47.493	5.0	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	49.773	0.5	98	0.00
65 PM	Chlorobenzene	50.000	49.625	0.8	97	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.746	-1.5	97	0.00
67 C	Ethyl Benzene	50.000	49.091	1.8#	96	0.00
68 T	m/p-Xylenes	100.000	98.735	1.3	97	0.00
69 T	o-Xylene	50.000	49.174	1.7	98	0.00
70 T	Styrene	50.000	50.570	-1.1	96	0.00
71 P	Bromoform	50.000	49.561	0.9	95	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	48.689	2.6	97	0.00
74 T	N-ethyl acetate	50.000	47.204	5.6	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.536	8.9	92	0.00
76 T	1,2,3-Trichloropropane	50.000	48.679	2.6	90	0.00
77 T	Bromobenzene	50.000	49.057	1.9	96	0.00
78 T	n-propylbenzene	50.000	49.166	1.7	97	0.00
79 T	2-Chlorotoluene	50.000	47.333	5.3	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.936	4.1	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.171	7.7	89	0.00
82 T	4-Chlorotoluene	50.000	47.614	4.8	96	0.00
83 T	tert-Butylbenzene	50.000	47.459	5.1	96	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.168	3.7	97	0.00
85 T	sec-Butylbenzene	50.000	48.519	3.0	97	0.00
86 T	p-Isopropyltoluene	50.000	49.090	1.8	97	0.00
87 T	1,3-Dichlorobenzene	50.000	50.280	-0.6	98	0.00
88 T	1,4-Dichlorobenzene	50.000	47.373	5.3	95	0.00
89 T	n-Butylbenzene	50.000	50.151	-0.3	96	0.00
90 T	Hexachloroethane	50.000	48.944	2.1	97	0.00
91 T	1,2-Dichlorobenzene	50.000	50.336	-0.7	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.630	8.7	87	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.490	-7.0	96	0.00
94 T	Hexachlorobutadiene	50.000	46.307	7.4	94	0.00
95 T	Naphthalene	50.000	49.142	1.7	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030221\
Data File : VN066112.D
Acq On : 2 Mar 2021 20:37
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 03 03:56:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 02 14:05:08 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.776	0.4	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6