

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN030320\
 Data File : VN060317.D
 Acq On : 3 Mar 2020 10:09
 Operator : JC/MD
 Sample : VN0303MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0303MBL01

Quant Time: Mar 04 01:35:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	193136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	314042	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	280941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	119307	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	121870	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
35) Dibromofluoromethane	7.56	113	89393	61.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	122.66%	
50) Toluene-d8	10.08	98	380270	51.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.08%	
62) 4-Bromofluorobenzene	12.40	95	132411	47.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030320\
Data File : VN060317.D
Acq On : 3 Mar 2020 10:09
Operator : JC/MD
Sample : VN0303MBL01
Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0303MBL01

Quant Time: Mar 04 01:35:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

