

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030320\
 Data File : VN060332.D
 Acq On : 3 Mar 2020 16:12
 Operator : JC/MD
 Sample : L1762-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-030220

Quant Time: Mar 04 02:40:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	194841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	317747	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	297894	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	139462	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124579	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
35) Dibromofluoromethane	7.57	113	90395	61.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.58%	
50) Toluene-d8	10.08	98	390915	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
62) 4-Bromofluorobenzene	12.40	95	145686	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
16) Acetone	3.79	43	5949	7.369	ug/l	100
20) Methylene Chloride	4.52	84	282414	127.384	ug/l	97
43) Isopropyl Acetate	8.14	43	37699	7.510	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030320\
Data File : VN060332.D
Acq On : 3 Mar 2020 16:12
Operator : JC/MD
Sample : L1762-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-030220

Quant Time: Mar 04 02:40:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

