

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030320\
 Data File : VN060333.D
 Acq On : 3 Mar 2020 16:35
 Operator : JC/MD
 Sample : L1762-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-06-030220

Quant Time: Mar 04 02:43:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205976	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	332471	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	316940	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	149257	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131291	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
35) Dibromofluoromethane	7.57	113	96560	62.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.14%	
50) Toluene-d8	10.08	98	413766	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
62) 4-Bromofluorobenzene	12.40	95	156034	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
Target Compounds						
16) Acetone	3.79	43	6361	7.454	ug/l	97
20) Methylene Chloride	4.52	84	508953	217.155	ug/l	98
43) Isopropyl Acetate	8.14	43	35393	6.738	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030320\
Data File : VN060333.D
Acq On : 3 Mar 2020 16:35
Operator : JC/MD
Sample : L1762-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-06-030220

Quant Time: Mar 04 02:43:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

