

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030321\
 Data File : VN066120.D
 Acq On : 3 Mar 2021 13:32
 Operator : JC/MD
 Sample : M1507-03ME
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41007ME

Quant Time: Mar 04 02:14:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 02 14:05:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.614	168	55606	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.540	114	102258	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	108623	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	52509	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.978	65	41902	49.52	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	99.04%	
35) Dibromofluoromethane	7.540	113	31818	49.13	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	98.26%	
50) Toluene-d8	10.052	98	128376	50.90	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	101.80%	
62) 4-Bromofluorobenzene	12.373	95	59878	61.64	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	123.28%	
Target Compounds						
						Qvalue
16) Acetone	3.785	43	239039	945.08	ug/l	96
18) Methyl Acetate	4.287	43	3324	6.16	ug/l #	81
39) Methylcyclohexane	9.036	83	5278	5.08	ug/l	98
51) 4-Methyl-2-Pentanone	9.949	43	6504	8.48	ug/l	97
52) Toluene	10.116	92	464657	254.46	ug/l	95
67) Ethyl Benzene	11.479	91	53923	13.20	ug/l	98
68) m/p-Xylenes	11.585	106	85247	54.01	ug/l	97
69) o-Xylene	11.913	106	27903	18.10	ug/l	98
73) Isopropylbenzene	12.219	105	4543	1.05	ug/l	95
78) n-propylbenzene	12.559	91	16772	3.59	ug/l	100
80) 1,3,5-Trimethylbenzene	12.701	105	24279	6.37	ug/l	93
84) 1,2,4-Trimethylbenzene	13.010	105	82605	22.17	ug/l	99
85) sec-Butylbenzene	13.141	105	5646	1.47	ug/l	84
86) p-Isopropyltoluene	13.257	119	4295	1.20	ug/l	97
89) n-Butylbenzene	13.585	91	8001	2.83	ug/l #	67
95) Naphthalene	15.100	128	24652	8.42	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030321\
Data File : VN066120.D
Acq On : 3 Mar 2021 13:32
Operator : JC/MD
Sample : M1507-03ME
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41007ME

Quant Time: Mar 04 02:14:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 02 14:05:08 2021
Response via : Initial Calibration

